

# TECHNICAL SPECIFICATION

## General Specification

Light Source	Xenon Flash Lamp
Detector	CCD [2,048 Pixels]
Wavelength Accuracy	±1 nm
Wavelength Range	190 – 1100 nm
Spectral Resolution	0.3 nm
Dimension	220 x 280 x 220 mm 8.6 x 11.0 x 8.6 in
Weight	5 kg
Operator Voltage	100 – 240 V, 50 – 60 Hz
Power Consumption	80 – 100 W
Warranty	1 Year
LCD Touch Panel	7.0 in

## Nano Volume Specificaiton

Absorbance Precision	1% at 100 ng/μl
Absorbance Range	0 – 200 Abs. (10 nm equivalent)
Detection Limit	2 ng/μl (dsDNA)
Maximum Concentration	15,000 ng/μl (dsDNA)
Measurement Time	5 sec
Minimum Sample Size	1 μl
Path Length	0.01 – 1.2 mm (Auto-ranging)

## Cuvette Specificaiton

Beam Height	8.5 mm
Absorbance Range	0.002 – 2.0 Abs.
Measurement Time	3 sec



### MicroDigital Co., Ltd.

CS-7th floor, 15, Pangyo-ro 228beon-gil,  
Bundang-gu, Seongnam-si, Gyeonggi-do,  
13487, Rep. of KOREA  
TEL: +82-31-701-2225 FAX: +82-31-702-2225  
<http://www.md-best.com>



# Nabi

## UV/Vis NANO SPECTROPHOTOMETER

[www.md-best.com](http://www.md-best.com)



MicroDigital Co., Ltd.

Nabi

UV/Vis NANO  
SPECTROPHOTOMETER

## Variety & Precision in Detection

- End Point & Kinetic Mode in Detection
- High Speed Spectrum Reading (5 Seconds)  
Accurate Detection (Less than  $\pm 0.002$ )
- Wide Range of Wavelength Detection with Cuvette
- Nano Volume Sample Measurement  
(Nucleic Acid / Protein)
- High Repeatability in Quantitative  
Measurement for Nucleic Acid

# Nabi UV/Vis NANO SPECTROPHOTOMETER



Nabi

UV/Vis NANO  
SPECTROPHOTOMETER

## Simple & Light in Design


- Compact Size and Light Weight
- Easy Operation with Minimized Steps  
for Measurements and Analysis
- 7.0 inch Touch LCD
- Quick Boot – No Lamp Warm-up Time
- Stand-Alone – Perform All Functions without PC

Nabi

UV/Vis NANO  
SPECTROPHOTOMETER

## Result Analysis

- Single Wavelength / Spectrum Selection
- End Point / Kinetic Mode Selection
- Quantitative Measurements for  
Nucleic Acid, Protein, and etc.
- Cell Concentration
- Simple Data Backup and Software  
Upgrade via USB Flash Drive
- Easy to Confirm the Result via Graphical Presentation



Time	Mode	Wavelength	Unit	Area
2019-08-26 02:14	END POINT	600	END POINT	0.000
2019-08-26 02:14	END POINT	600	END POINT	0.012
2019-08-26 02:15	END POINT	600	END POINT	0.210
2019-08-26 02:15	END POINT	600	END POINT	0.171
2019-08-26 02:15	END POINT	600	END POINT	0.000
2019-08-26 02:15	END POINT	600	END POINT	0.200



Time	Mode	Wavelength	Unit	Area
2019-08-26 02:14	END POINT	600	END POINT	0.000
2019-08-26 02:14	END POINT	600	END POINT	0.012
2019-08-26 02:15	END POINT	600	END POINT	0.210
2019-08-26 02:15	END POINT	600	END POINT	0.171
2019-08-26 02:15	END POINT	600	END POINT	0.000
2019-08-26 02:15	END POINT	600	END POINT	0.200